

## 100 MHz Analog Oscilloscope

- Dual time base oscilloscope (2 channel)
- 5mV/division sensitivity
- Sweeps to 5ns/division
- 23 calibrated ranges, main time base

model

- Signal delay line
- 15 kV accelerating voltage
- Channel 2 output

	_
Specification	ns
vertical amplifiers (Ci	H 1 and CH 2)
Sensitivity	5mV/div to 5 V/div. 1 mV/div to 1V/div (at X5 MAG)
Attenuator	10 calibrated steps in 1-2-5 sequence.
	Vernier control provides fully adjustable sensitivity
	between steps, adjustment range 1/1 to 1/2.5
Accuracy	±3% (±5% at X5 MAG)
Input Impedance	1ΜΩ +3%
Input Capacitance	25 pF ±10pF
Frequency Response	DC: DC to 100 MHz (-3 dB), on 10mV/Div setting
X5 MAG	DC to 20 MHz (-3 dB)
AC	10Hz to 100 MHz (-3 dB)
Rise Time	3.5 ns (Overshoot <u>&lt;</u> 5%)
Signal Delay Time	Variable
Square Wave Characteristics	Overshoot less than 5%, 10 mV/div range
	Other ranges within 5% additional
Maximum Input Voltage	400V (DC + AC peak)
/ertical amplifiers	
Operating Modes	CH 1, CH 2, Dual, Add
Delay Time Between Channels	Within 1 ns between CH 1 and CH 2
Crosstalk	30:1 at 100 kHz
SWEEP SYSTEM Operating Modes	
A	A sweep
В	Delayed B sweep
B TRIGGERED	B sweep triggered after delay
A Time Base	b sweep triggered after delay
Sweep Mode	Auto, normal
Sweep Node Sweep Time:	5s to 20ns/div., 23 steps in 1-2-5 sequence
	with variable control
Accuracy	+ 3%
Hold Off Time	Continuously variable. Adjustment range from
Hold Oil Time	normal to 1.5 times the sweep time
B Time Base	normal to 1.5 times the sweep time
Delay Method	Continuous delay. Triggered delay
Sweep Time	20ns. to 0.5s/div., 23 steps in 1-2-5 sequence
Accuracy	± 3%
Delay Time	Start point: 0.5 div to + 0.3 div.
Doidy Time	End point: $0.3 \text{ div to } + 0.3 \text{ div.}$
Delay Jitter	Within 1/10,000 of full scale sweep time
Delay sitter	Within 1710,000 or full scale sweep time
TRIGGERING	
A Trigger	
Source	CH 1, CH 2, LINE, EXT
Sensitivity	30Hz to 110MHz
TV-V	20Hz - 30kHz
TV-H	3kHz - 100kHz
Slope	+ Or -
B Trigger	The A trigger is also the B trigger
D 1119901	The A digger is also the D trigger

2190B
1mΩ, 30pF
300V (DC + AC peak)
X Axis = CH 1. Y Axis = CH 2
5 mV/div to 5 V/div, CH 1 and CH 2
$\pm$ 3% calibrated position, $\pm$ 6% using x10 MAG
DC to 2 MHz (-3dB)
Approx. 100mV/div open circuit
Approx. 50 mV/div into 50Ω
50 Hz to 30 MHz.
approx. 50Ω
Rectangular with integral graticule
8 x 10 div (1 div = 1 cm)
15kV
P31
None
Electrical, front panel adjustable

## Other Specifications Sensitivity: 3 V or greater, TTL level. (Intensity Modulation) Negative polarity increases brightness Input Impedance 15 kΩ Usable Freq. Range DC to 3.5 MHz 20 V (DC + AC peak) Maximum Input Voltage CAL/Probe Compensation Waveform Positive going squareware Output Voltage $0.5~\text{V p-p}~\pm3\%$ Frequency Approx. 1kHz Duty Cycle $50~\pm~5\%$ Power Requirements 100/120/220/240/ VAC ±10%, 50/60 Hz, approximately 55 W Dimensions (HxWxD) 12.76 x 15.68 x 5.2" (324 x 398 x 132 mm) Weight 18.7 lbs (8.5 kg) **ENVIRONMENT** 50° to 95°F (10° to 35°C), 85% maximum RH Within Specified Accuracy Full Operation 32 $^{\circ}$ to 104 $^{\circ}\text{F}$ (0 $^{\circ}$ to +40 $^{\circ}\text{C}$ ), 85% maximum RH

## **Accessories**

Storage

Three Year Warranty

SUPPLIED: Instruction Manual, Two PR-37A x1/x10/Ref. Probes or equivalent, AC Power Cord, Spare Fuse

-4° to 158°F (-20°to +70°C)

OPTIONAL: PR-32A Demodulator Probe, PR-46A x10 Probe, PR-37A x1/x10/REF. Probe, PR-100A x100 Probe, PR-55 High Voltage x1000 Probe, LC-210A Carrying Case

