

# CONTRACTOR'S CHOICE - TEST LEAD KIT

**DESIGNED SPECIFICALLY FOR ELECTRICAL,  
HVAC, AND SECURITY APPLICATIONS**

**P/N 905232**



- **Compatible with all Industry Standard 4mm Instrument (DMM) Sockets**
- **RATED 1000V CAT III/600V CAT IV, 20 Amps**

**Includes:**

- 1 Pair Modular, Double Insulated, Silicone Leads, 48"
  - 1 Pair Extra Sharp, Stainless Steel Probes
  - 1 Pair Fully Insulated Crocodile Clips
  - 1 Pair Electric Receptacle Test Adaptors
  - 1 Pair Sprung Hooks
  - 1 Pair Probe Tip Insulation Extenders
- Packed In a Tool Roll for Convenient Storage**



**SILVERTRONIC LTD**


Title: Test Lead Set

Material: N/A

This drawing is not to be disclosed to a third party without the written authority from the company ©

Rev	Details	Date	Drawn	Part No:
4				129401
3				
2	Packaging No. changed.	V.N	16/02/04	Project No: N/A
1	First issue	#	#	
-	Pre-Production issue	03/04/03	Rich	

OR Plug: 125076B




OF Plug: 125070B



119054B D/S Cable cut 1250mm or (102523)

OR Plug: 125076R




OF Plug: 125070R



119054R D/S Cable cut 1250mm or (102523)


ISO metric view of USA Mains tester




USA Mains Tester: 121328B




USA Mains Tester: 121328R




1000V Croc. Clip: 126024B




OEM Style Prod: 125072B




OEM Style Prod: 125072R




Sprung Hook: 121026B



1000V Croc. Clip: 126024R



Sprung Hook: 121026R



Assembly Jig Numbers

#####=J###

#####=J###

#####=J###

#####=J###

Printing as illustrated

Packaging: 119512 Pouch

Tolerances:  
Cable Lengths +/- 5%

Do Not Scale All Dimensions In MM

## DECLARATION OF CONFORMITY

**SILVERTRONIC Ltd. confirm that the products listed below have been Designed, Manufactured and Tested to, and therefore are fully compliant to;**  
**IEC1010-1 1990 inc. Amendment 2-1995 &**  
**IEC1010-2-031 First Edition-1993**  
**and thus meet THE LOW VOLTAGE DIRECTIVE (LVD)**  
**73/23/EEC and amending directive 93/68/EEC**

PRODUCT TYPE / DESCRIPTION				Test Lead Set	
PART No.				129401	
IEC1010 RATING	VOLTAGE	1000	Volts	INSULATION CATEGORY	III ( 3 )

MAXIMUM RATED CURRENT	*20/30	<b>Amps</b>
POLLUTION DEGREE	2	
INSULATION CATEGORY	DOUBLE/REINFORCED	

INCLUDES FOLLOWING ASSEMBLIES / COMPONENTS SUBJECT TO INDIVIDUAL TEST.	DESCRIPTION	PART No.
	*OEM Full Safety Plug To OEM Right Angle Plug Lead	123815
	OEM Style Prod	125072
	1000V Croc. Clip	126024
	Sprung Hook	121026
	U.S.A. Mains Tester	121328



All Testing carried out by Third Party Testhouse  
**CRANAGE EMC TESTING Ltd.**  
( LVD TESTING DEPARTMENT )



www.silvertronic.com  
e-mail: sales@silvertronic.com

Registration No. 1993292

**North American Office**  
4405 Sharon Copley Road,  
Medina, Ohio, 44256.  
Toll Free: (866) 556 3308  
Facsimile: (330) 722 2281