

# GSM28 Medical

ONDO

28 Watt Global Performance Switchers

# **GLOBAL PERFORMANCE SWITCHERS**

# **FEATURES:**

- Industry's smallest 28 W medically approved switcher
- Compact size (4.00" x 2.59" x 0.92")
- Wide-range ac input: 85-264 Vac
- Less than 25 μA leakage current @ 120 Vac
- Approved to UL2601-1, EN60601-1
- EMI to FCC, CISPR 11 Class B/IEC601-1-2
- Overvoltage protection standard
- RoHS Compliant Models Available (G suffix)
- CE marked to LVD





# SPECIFICATIONS

Ac Input
----------

# 85-264 Vac, 47-63 Hz single phase.

# Input Current

Maximum input current at 120 Vac, 60 Hz with full rated output load is 0.85 A. Hold-up Time

15 ms minimum from loss of ac input at full load, nominal line (120 Vac).

## **Output Power**

Normal continuous output power is 28 W, 32 W peak for 60 s maximum duration, 10% duty cycle. Factory set to begin power limiting at approximately 35 W.

#### **Overload Protection**

Fully protected against short circuit and output overload. Short circuit protection is cycling type power limit.

#### **Output Noise**

0.5% rms, 1% pk-pk, 20 MHz bandwidth, differential mode. Measured with scope probe directly across output terminals of the power supply with load terminated with 0.1  $\,\mu F$  capacitor.

# Overvoltage Protection

Built in with firing point set per ratings table. OVP firing reduces voltage to less than 50% of nominal voltage in 50 ms.

## Voltage Adjustment

Factory set with fixed resistors to maximize reliability.

#### Efficiency

70% minimum for the 5.1 V model at full rated load, nominal input voltage. Efficiency increases as output voltage increases.

## Input Protection

Internal ac fuse provided on all units. Designed to open only if a catastrophic failure occurs in the unit.

# Inrush Current

Inrush limited by internal thermistor. The inrush at 230 Vac, averaged over the first ac half-cycle under cold start conditions will not exceed 32 A.

Minimum Load Requirement 5% of full load rating

Transient Response 3.5% max. dev. 50% load step at 0.2 A/µs. Recovery to 0.5% within 500  $\mu s.$ 

Temperature Coefficient 0.03%/°C typical on all outputs.

# Environmental

Designed for 0 to 50°C operation at full rated output power; derate output current and total output power by 2.5% per °C above 50°C.

Medical EMI/EMC Compliance

All models include built-in EMI filtering to meet the following emissions requirements:

EMI SPECIFICATIONS	COMPLIANCE LEVEL
Conducted Emissions	EN55011 Class B; FCC Class B
Static Discharge	EN61000-4-2, 6 kV contact, 8 kV air
RF Field Susceptibility	EN61000-4-3, 3 V/meter
Fast Transients/Bursts	EN61000-4-4, 2 kV, 5 kHz
Surge Susceptibility	EN61000-4-5, 1 kV diff., 2 kV com.
Line Frequency Harmonics	EN61000-3-2 Class A

## Earth Leakage Current

Leakage current measured in the Gnd wire connection when measured per EN60601-1 or UL2601-1 is as follows:

Medical Model	Normal Leakage	SingleFault Leakage	Test Voltage	Test Method
GSM28	25μΑ	45 μΑ	132 Vca/60 Hz	UL2601-1
GSM28	50 µA	90 µA	264 Vca/50 Hz	IEC60601-1


#### Medical Medical Safety

SL Power Electronics Corp. declares under our sole responsibility that all GSM models are in conformity with the applicable requirements of UL2601-1 Patient Care Equipment, CSA-C22.2 No. 234 (with additional tests to C22.2 No. 601.1 per T.I.L. CA-08), EN60601-1.

Medical Model	Output	Current	Load Regulation	Initial Setpoint Tolerance	OVP Setpoint	Ripple and Noise
GSM28-5	5.1 V	5.5 A	0.75%	2.5%	$6.2\pm0.6V$	1.4%
GSM28-12	12 V	2.3 A	0.75%	2.5%	14 ± 1.0 V	1%
GSM28-15	15 V	1.9 A	0.75%	2.5%	18.5 ± 1.5 V	1%
GSM28-24	24 V	1.2 A	0.75%	2.5%	28 ± 2.5 V	1%
GSM28-28	28 V	1.0 A	0.75%	2.5%	$34 \pm 2.8$ V	1%

\* Add "G" suffix to model number for RoHS compliant model.

# GSM28 MECHANICAL SPECIFICATIONS



ENVIRONMENTAL SPECIFICATIONS	OPERATING	NON-OPERATING
Temperature (A)	See individual specs	-40 to +85°C
Humidity (A)	0 to 95% RH	0 to 95% RH
Shock (B)	20 g <sub>pk</sub>	40 g <sub>pk</sub>
Altitude	-500 to 10,000 ft	-500 to 40,000 ft
Vibration (C)	1.5 g <sub>rms'</sub> 0.003 g²/Hz	5 g <sub>rms'</sub> 0.026 g²/Hz

A. Units should be allowed to warm up/operate under non-condensing conditions before application of power.

B. Shock testing—half-sinusoidal, 10  $\pm$  3 ms duration,  $\pm$  direction, 3 orthogonal axes, total 6 shocks.

C. , Random vibration—10 to 2000Hz, 6dB/octave roll-off from 350 to 2000Hz 3 orthogonal axes. Tested for 10 min./axis operating and 1 hr./axis non-operating.